

TEST, MEASUREMENT, AND DIAGNOSTIC EQUIPMENT (TMDE)

Proponent For Inspection: DOL

Point of Contact: _____

Unit Inspected: _____

Date of Inspection: _____

Unit Representative: _____

Unit Phone No: _____

Inspector's Name: _____

Inspector's Phone No: _____

Unit Overall Rating: T P U

REFERENCES: a. AR 750-1, Jan 06 b. AR 750-43, Dec 04 c. TB 43-180, Current IETM (EM 0022)
d. TB 750-25, March 97 e. DA Pam 750-8, Aug 05 f. DA Pam 750-35, Aug 94
g. TSC-FLW SOP Number 738-1, March 05
h. Current applicable TMs for the equipment being inspected.

STANDARDS: "T"= 90% success rate of evaluated tasks with no failed critical tasks. "P"= 70% success rate of evaluated tasks with no failed critical tasks. "U"= less than 70% success rate of evaluated tasks or one failed critical task

INSPECTION CRITERIA:	LEVEL	GO	NO	REMARKS
1. Has the commander appointed in writing a designated representative as TMDE Support Coordinator to serve as the central point of contact between the unit and the TMDE Support Activity (TSA)? (AR 750-43, Para 6-4; TB 750-25, Para 3-8b, Para 3-10a a (1))	BN			
2. As a minimum, does the TMDE coordinator have the following publications: AR 750-1, AR 750-43, TB 750-25, TB 43-180, DA Pam 750-8, and the TSA External SOP? (TB 750-25, Para 3-10a(3))	BN			
3. CRITICAL: Has the organization met the minimum DA goal of having 95% or better of the TMDE inventory calibrated and in repaired condition and available to the user, with no more than a 2% delinquency rate? (AR 750-43, Para 6-1a & b)	BN			
4. Does the TMDE coordinator have a current TMDE Master List or Instrument Master Record File (IMRF) provided by the TSA quarterly, identifying all instruments belonging to the user? (AR 750-43, Para 6-12, Para 6-15 e (3); TB 750-25, Para 3-2b(2) & c, Para 3-10a(4))	BN			
5. Does the TMDE coordinator have a current Delinquent Items List, provided by the TSA monthly, identifying TMDE not presented for calibration within the scheduled timeframe? Is this list reviewed and is corrective action taken? (AR 750-43, Appendix B Para B-4h; TB 750-25, Para 3-2b(2), Para 3-10a(6), Para 3-10b(7))	BN			
6. Does the TMDE coordinator have a current Projected Items List (PIL), provided by the TSA at least monthly (minimum 30-day notice), identifying each instrument in the IRMF due for C&RS within that month? Is this list reviewed for accuracy? (TB 750-25, Para 3-2b, Para 3-10a(5), Para 3-10b(6))	BN			
7. Does TMDE in use have a current DA Label 80 or DA Label 163 affixed and properly annotated? (AR 750-43, Para 6-14; TB 750-25, Para 3-10b(12))	BN			
8. Has an operator/unit level maintenance program been established for TMDE (including CNR labeled TMDE) IAW equipment maintenance manuals and procedures in TB 750-25 Para 3-10? (TB 750-25, Para 3-10b(10) & (11))	BN			
9. Has an SOP been developed and implemented for the identification and control of TMDE requiring calibration and repair support? (TB 750-25, Para 3-10a(2))	BN			
REMARKS AND FINAL COMMENTS:				